

F. LIST GmbH
Attn. Mr. Jagersberger
List-Straße 1
2842 Edlitz-Thomasberg
AUSTRIA

Unser Zeichen	Ihr Zeichen	Bearbeiter	Tel.DW	Fax DW	Datum
AOT101-112/07-16	-	N. OTIN M. MARKUS	1630 1636	1666	25.05.2017

Subject: Acceptance of the F. LIST GmbH (LIST) Flammability Test Cabinet
by Austro Control GmbH (ACG)

Dear Mr. Jagersberger

Based on F.LISTs request dated 11.04.2016, in respect of the addition of test operators and one test witness to the authorized personal of the Flammability Test Laboratory and the update of several internal documents, ACG herewith confirms that the Flammability Test Cabinet, located at F. LIST GmbH in Edlitz-Thomasberg (Austria) still fulfils all the requirements stated in the Aircraft Materials Fire Test Handbook (DOT/FAA/AR-00/12, Chapter 1 and 3).

Based on the following items, ACG accepts the use of this test cabinet for flammability tests:

- Comparative Round Robin tests accomplished by Isovolta, Aeroworks and LIST (see Doc. Nr. MFP-0606 Rev A dated 21. Oct 2013 – Round Robin Comparative Test Plan/Report),
- Physical Inspections, witnessing and assessments performed by ACG
- Report demonstrating the compliance of the Test Cabinet against FAA Fire Test Handbook and Flammability Test Procedures, (see Doc. Nr. AWD-0608 "Operating Manual" Rev. C or later approved revision)

The scope of flammability tests accepted to be performed in this cabinet by the present letter is restricted to:

Vertical Flammability Tests
as addressed in CS/FAR/JAR 23, 25, 27 und 29, §853 and §855

Horizontal Flammability Tests
as addressed in CS/FAR/JAR 23, 25, 27 und 29, §853 and §855

and as defined under CS/FAR/JAR 23/25 Appendix F, Fire Test Handbook DOT/FAA/AR-00/12, Chapter 1 and 3, and FAA AC 23-2A.

Seite 1 von 2

The personnel involved in the testing (operator and test witness) must be qualified and authorized according to "LIST Bildungsmatrix Production Organisation" Doc. Nr. QMD-0524. Rev. V dated 12.08.2013 or later approved revision and the list of personnel is recorded in "Authorized Personnel" Doc. Nr. AWD-1143. Rev. B or later approved revision. Delegation of Conformity Inspections and Test Witnessing activities from ACG to LIST (when requested) has to be notified according to the "Provisional Delegation Guidelines" Doc. Nr. AWD-0109 Rev. A or later approved revision

This acceptance remains valid until revocation. Any change in location, scope of test, personal or anything related to the flammability tests and associated test apparatus has to be notified to ACG, revised in the aforementioned reports and approved by ACG. This acceptance remains valid for any future ACG approved revisions of the aforementioned documents. ACG can perform, if judged necessary, inspections both periodic and ad hoc and request comparative tests of the Test Cabinet.

For Austro Control GmbH




for _____
Wilhelm Jagritsch
Head Airworthiness and Certification